

Test Verification of Conformity

Certificate No.: CTE25040100

Date: 2025-06-20

In accordance with the following Applicable Directives:

2014/30/EU

Electromagnetic Compatibility

The equipment, as described herewith, was tested pursuant to applicable test procedure and complies with the requirements of:

EN IEC 61326-1: 2021 EN IEC 61326-2-4: 2021

EN 61000-4-2: 2009 EN 61000-4-4: 2012

EN 61000-4-6: 2014 EN 61000-4-8: 2010

EN 61000-4-11: 2004+A1: 2017

EN 61000-4-3: 2006+A1: 2008+A2: 2010 EN 61000-4-5: 2014+A1: 2017

The test results are traceable to the international or national standards.

Applicant: **CET Electric Technology Inc.**

Floor 33-35, Building #6, Shenzhen International Innovation Valley, Dashi First Road,
Nanshan District, Shenzhen, Guangdong, P.R. China 518055

Manufacturer: **CET Electric Technology Inc.**

2,3,4,5 floors of Building 4, Beishan Industrial Area, Beishan Street, Yan Tian ,
Shenzhen, Guangdong, P.R.China

EUT Name: **Medical Insulation Monitoring Device**

Model number: **IMH100-P11AAAE**

Listed Model(s): **IMH100-ABCDEF (detailed model differentiation see page 5 of report)**

Laboratory: **Shenzhen Huatongwei International Inspection Co., Ltd.**

Huatongwei 101, No.006, Keji south 12th Road, High-tech zone community, Yuehai Street, Nanshan District,
Shenzhen, Guangdong, China

Tel: 86-755-26715499

E-mail: cs@szhtw.com.cn

Website: Http://www.szhtw.com.cn

Note:

The certification is only valid for the equipment and configuration described, in conjunction with the test data detailed above.

The CE mark as shown beside can be used, under the responsibility of the manufacturer, after completion of an EC Directive of
Conformity and compliance with all relevant EC Directive.

For and on behalf of

Shenzhen Huatongwei International Inspection Co., Ltd.

Authorized by:

Tiger Jiang

